

VFC400 Installation



Installation

Your kit includes:

- VFC400 Data Logger
- Stainless Steel Probe encased in glycol
- Acrylic stand for probe and mounting equipment for logger
- Adhesive backed zip tie mounts and zip ties for securing cable
- Spare battery
- 2 year NIST traceable Certificate of Calibration compliant to ISO 17025: 2005

Calibration Certificate

Certificate ID:

Customer Information		Equipment Information	
Customer:	Control Solutions, Inc.	Equipment ID:	V17-4000
Address:	35851 Industrial Way, Suite D	Model No:	VFC400
City, State Zip:	St. Helens, OR 97051	Manufacturer:	LOG TAG
Contact:		Serial No:	1860922670
Phone:	503-443-5411		

Calibration Summary	
Temp:	70F +/- 10
Humidity:	50% +/-20
Technician:	Nicole H

As Found	Result
In Tolerance	Pass

As Found



Installation



1

Place acrylic stand and probe vial near the middle of the fridge/freezer

2

Route the cable under the wire rack and secure with zip tie

3

Route the cable toward the wall of the hinge side and secure with zip tie



Installation

- Route the cable toward the front of the fridge/freezer on the hinge side and secure

**Place the glycol bottle in the fridge/freezer for at least 1.5 hours prior to starting your logger to allow the solution to reach the appropriate temperature.



Installation



- Adhere the mounting bracket to the side or front of your fridge/freezer
- Place the logger into the mounting bracket and plug the sensor wire into the logger (left side)
- Approx. 6 inches underneath the logger, adhere cable tie bracket and secure the cable with a Zip tie. Leave enough slack in the cable so you can plug and unplug the VFC400 easily

Questions about installation?

Visit our [support page](#)

Send us an email
support@vfcdataloggers.com

Jump to:

[VFC400 Installation](#)

[VFC400 Software Download](#)

[VFC400 Software Set Up](#)

[VFC400 Logger Configuration](#)

[Data Logger Operation](#)

